


# EXHIBIT 4

 LANGFORD IC SYSTEMS, INC.	<b>MANZI MACH 1 – INSTRUMENT CLEANER - PROCESSOR</b> <b>Pre-Market Notification 510(k) # K060458</b> <b>Section II Tab 10</b> <b>510(k) Summary</b>	<b>Page 1 of 4</b>
---	--	--------------------

## **Applicant's Name, Address, Telephone, Fax, Contact Person**

Langford IC Systems, Inc.  
310 S. Williams Blvd, Suite 270  
Tucson, Arizona 85711  
Tel: 520-745-6201  
Fax: 520-745-6286  
E-mail: [terrylangford@licsystems.com](mailto:terrylangford@licsystems.com)

## **Contact Person**

Terry Langford  
President and CEO, Langford IC Systems, Inc.  
Tel: 520-745-6201  
Fax: 520-745-6286  
E-mail: [terrylangford@licsystems.com](mailto:terrylangford@licsystems.com)

## **1.0 CLASSIFICATION, COMMON OR USUAL NAME, DEVICE NAME**

- a) Classification Name: Pending - Class II  
Common / Usual Name: Endoscope and accessories  
Device Classification: 21 CFR § 876.1500, –Endoscope and accessories  
Proprietary Name: Manzi Mach 1 Instrument Cleaner- Processor System with MS10 High Level Disinfectant
- b) Classification Name: Pending - Class II  
Common / Usual Name: Endoscope and accessories  
Device Classification: 21 CFR § 876.1500,  
Proprietary Name: Manzi MS10

## **2.0 PREDICATE DEVICE**

- a) System 83 Plus™ Washer-Disinfectant, K983017  
Manzi Cleaner System, K043314 (Washer)
- b) Steris 20 Sterilant, K875280

## **3.0 INDICATIONS FOR USE**

The Manzi Mach 1 Instrument Cleaner Processor System is indicated for use with the High Level Disinfectant MS10 concentrate (MEC 0.49% PAA, minimum contact temperature of 120°F for a contact time of 15 minutes) for cleaning and high level disinfecting flexible bronchoscopes used in health care settings by health care workers.


## **4.0 DESCRIPTION OF THE DEVICE**

The Manzi Mach 1 Instrument Cleaner-Processor System consists of a Manzi Mach 1 Instrument Cleaner-Processor; a proprietary Manzi germicide, MS10; and a proprietary Manzi Detergent, MD10.

The Manzi Mach 1 Instrument Cleaner-Processor is a self-contained stand-alone system of hardware and software designed to clean and provide high level disinfection of bronchoscopes using the MD10 detergent, the MS10 germicide, and a patented push-pull agitation system. The push-pull agitation system effectively scrubs the interior and exterior surfaces of the bronchoscope without the use of special connectors. The scope is placed in a processing chamber where it is exposed to a push-pull

10/12/06

**EXHIBIT 4**

 LANGFORD IC SYSTEMS, INC	<b>MANZI MACH 1 – INSTRUMENT CLEANER - PROCESSOR</b> <b>Pre-Market Notification 510(k) # K060458</b> <b>Section II Tab 10</b> <b>510(k) Summary</b>	Page 2 of 4
--	--	-------------

agitation cleaning cycle followed by two hot water rinses, a push-pull agitation disinfection cycle that provides high level disinfection (spore log reduction of  $\geq 10^6$  microorganisms with no CFUs) of the device, and a final rinse with an ozonated sanitized water rinse.

The hardware for the Manzi Mach 1 Instrument Cleaner-Processor consists of a stainless steel processing chamber, a push-pull agitation pump, an ozonator, and a variety of components that are mounted in a movable covered frame. The cleaner-processor system utilizes accessories such as disposable water filters, reusable bronchoscope trays, and printer paper.

The Manzi Mach 1 Instrument Cleaner-Processor is designed to: (1) be used in accordance with the reprocessing instructions provided in the operator's manual of the instruments being processed, and (2) facilitate the health care facility's compliance with reprocessing guidelines published by SGNA, APIC, AORN, ASGE, CDC, and other professional organizations.

MD10 is a low foaming enzyme chemical detergent packaged in single use containers for attachment to the Manzi Cleaner. MD10 is intended to be used with the Manzi Instrument Cleaner-Processor.

MS10 is a peracetic acid based liquid chemical germicide. MS10 is intended to be used with the Manzi Instrument Cleaner-Processor.

## 5.0 SUMMARY OF NONCLINICAL TESTS for the MANZI MACH 1 INSTRUMENT CLEANER-PROCESSOR

### 5.1 Qualification Testing – FDA Guidance

The Manzi Mach 1 Instrument Cleaner-Processor System was tested and found to conform with the requirements of the “**Guidance on Premarket Notification [510(k)] Submissions for Automated Endoscope Washers, Washer / Disinfectors, and Disinfectors Intended for Use in Health Care Facilities**”, dated August, 1993”. The table below identifies the qualifications performed and the results obtained:

Requirement		Results
G..	Performance Data	
1	Process Parameter Tests	Passed
2	Simulated Use Tests	Passed
2.c.	Effectiveness Tests	
2.c.(1)	Cleaning Efficacy	Passed
2.c.(2)	Disinfection Efficacy	Passed
2.c.(3)	Rinsing Efficacy	Passed
2.c.(4)	Other Tests	Passed
2.c.(5)	Combined Process	Passed
3.	In – Use Tests	Passed
H.	Software Documentation	Passed
I.	Toxicological Evaluation of Residues	Passed

### 5.2 Qualification Testing – EU Guidance

10/12/06

The Manzi Mach 1 Instrument Cleaner-Processor System was also tested and found to conform with the requirements of the Draft prEN ISO 15883-1: 2003, Washer-disinfectors – Part 1: General Requirements, Definitions and Tests and Draft prEN ISO 15883-4: 2001, Washer-disinfectors – Part 4: Requirements and Tests for Washer-Disinfectors Employing Chemical Disinfection for Thermo-Labile Endoscopes as identified in the Table below.

<b>prEN ISO 15883-1: 2003 Requirement</b>		<b>Results</b>
6.10 Annex B Annex E	Cleaning Efficacy – Scope Ninhydrin Horse serum- prEN 15883-4 Annex B.1.1	Passed
	Cleaning Efficacy – Surrogate Ninhydrin Horse serum- prEN 15883-4 Annex B.1.1 Surrogate - prEN 15883-4	Passed
6.11 Annex D	Disinfection Efficacy – Scope Sheep blood - prEN 15883-4 Annex D	Passed
	Disinfection Efficacy – Surrogate Sheep blood - prEN 15883-4 Annex D	Passed

### 5.3 Qualification Testing – Langford IC Systems (LIC) Requirements

The Manzi Mach 1 Instrument Cleaner-Processor System was also tested and found to conform to the LIC requirements identified in the table below.

<b>LIC Cleaning Requirement</b>	<b>Results</b>
Cleaning Efficacy: Reduction of protein loading of scopes contaminated with a Protein Laden Soil to Remaining Protein levels of $< 6.4 \mu\text{g}/\text{cm}^2$ (Ref: AAMI TIR30: 2003, A Compendium of Processes, Materials, Test Methods, and Acceptance Criteria for Cleaning Reusable Medical Devices.).	Passed Remaining Protein levels of $< 4.0 \mu\text{g}/\text{cm}^2$

<b>LIC High Level Disinfection Requirement</b>	<b>Results</b>
High Level Disinfection Efficacy: Reduction of $\geq 10^6$ microbial loading of scopes with no colony forming units (CFUs)	Passed $\geq 6$ spore log reduction with no CFUs

<b>LIC Microbiological Efficacy Tests</b>		
<b>Test Method</b>	<b>Test Organisms</b>	<b>Results</b>
Simulated Use Test (15 min.) ISO 15883-4 Surrogates Sheep's blood soil	Bacillus subtilis Mycobacterium terrae Candida albicans Enterococcus faecium Styphlococcus aureus	$> 6$ spore log reduction; no Colony Forming Units (CFU)
Simulated Use Test (15 min.) Olympus Bronchoscopes Sheep's blood soil	Bacillus subtilis	$> 6$ spore log reduction; no CFU
Simulated Use Test Ozonated Water System	Candida albicans, Staphylococcus aureus, Bacillus subtilis	$> 6$ spore log reduction; no CFU

LIC Sanitized Ozonated Water System Requirement	Results
Final Rinse System Efficacy: Reduction of $\geq 10^6$ microbial loading of scopes with no colony forming units (CFUs)	Passed $\geq 6$ spore log reduction with no CFUs

## 6.0 SUMMARY OF NONCLINICAL TESTS for the MANZI MS10

### 5.1 Qualification Testing – FDA Guidance

The Manzi MS10 germicide was tested to and met the requirements of the current edition of “Guidance for Industry and FDA Reviewers, Content and Format of Premarket Notification [510(k)] Submissions for Liquid Chemical Sterilants / High Level Disinfectants”, dated January 3, 2000. The table below identifies the qualifications performed and the results obtained:

Requirement	Results
5.4 Potency Test	Passed
5.5 Simulated Use Tests	Passed
5.6 In-Use Tests	Passed
6.0 Biocompatibility	Passed

Microbiological Efficacy Summary		
Test Method	Test Organisms	Results
Sporicidal Activity of Sterilants; AOAC Official Method 966.04	Bacillus subtilis Clostridium sporogenes	> 6 spore log reduction; No CFUs MS10 is sporicidal
Fungicidal Activity of Sterilants; AOAC Official Method 955.17	Trichophyton mentagrophytes	MS10 is fungicidal
Use-Dilution Method; AOAC Official Method 955.14, 955.15, 964.02	Salmonella choleraesuis Staphylococcus aureus Pseudomonas aeruginosa	MS10 is bactericidal
Virucidal Testing	Poliovirus Type 1	MS10 is virucidal
Quantitative Tuberculocidal Test	Mycobacterium bovis	MS10 is tuberculocidal

## 7.0 OVERALL PERFORMANCE CONCLUSIONS

The studies demonstrate that the Manzi Mach 1 Instrument Cleaner-Processor System is safe and effective for the cleaning and high level disinfection of bronchoscopes within the stated indications for use for the Manzi Mach 1 Instrument Cleaner-Processor, the Manzi MS10 germicide, and the Manzi Detergent, MD10, and establishes substantial equivalence of the Manzi Mach 1 Instrument Cleaner-Processor System to the predicate devices identified in Section 2.0 above.